



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 078389	PRINTER SN 071.3383.132	DATE OF INSPECTION 09/28/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 4525 Downs Drive St. Joseph, MO 64507		TIME OF INSPECTION 8:53 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☒ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

☒ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

☒ PRINTER WORKING PROPERLY

☒ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☒ SIMULATOR SOLUTION

☐ COMPRESSED ETHANOL-GAS MIXTURE

☒ STANDARD SUPPLIER RepCo Marketing Inc. LOT # 13002 EXP. DATE 06/15/2015

☒ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN 3506 SIMULATOR EXP DATE 12/06/2014

☒ CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .101

TEST 2 .101

TEST 3 .101

☒ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument is operating within the limits set forth by the Missouri Department of Health and Senior Services. Solution provided by RepCo Marketing Inc. .100 solution, Lot# 13002 expires on 06/19/2015.

INSPECTING OFFICER

SIGNATURE

James Bench

PRINT NAME

James Bench

TYPE II PERMIT NUMBER/EXPIRATION DATE

240032 02/11/2016

TELEPHONE NUMBER

(816) 271-4438

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

AS IN Serial No. 978283
Serial No. 500

1937 FORD 4847

Year Make Type 2101

Mr. Blank

03/28/14 2014 .000
California Chgo
12 03/28/14 2014 .101

Subject Name

Subject I.R.

AS IN Serial No. 978283
Serial No. 500

1937 FORD 4847

Year Make Type 2101

Mr. Blank

03/28/14 2014 .000
California Chgo
12 03/28/14 2014 .101

Subject Name

Subject I.R.

AS IN Serial No. 978283
Serial No. 500

1937 FORD 4847

Year Make Type 2101

Mr. Blank

03/28/14 2014 .000
California Chgo
12 03/28/14 2014 .101

Subject Name

Subject I.R.

AS IN Serial No. 978283
Serial No. 500

1937 FORD 4847

Year Make Type 2101

Mr. Blank

03/28/14 2014 .000
California Chgo
12 03/28/14 2014 .101

Subject Name

Subject I.R.

Serial No. 2117
John Blank

4825 Downs Dr

St. Joseph, Mo 64507

Serial No. 2117
John Blank

4825 Downs Drive

St. Joseph, Mo 64507

Serial No. 2117
John Blank

4825 Downs Drive

St. Joseph, Mo 64507

Serial No. 2117
John Blank

4825 Downs Drive

St. Joseph, Mo 64507



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JAMES E BENCH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.


DATE 2/11/2014

NUMBER 240032



EXPIRES 2/11/2016

MO 580-0771 (6-10)


DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

	STATE OF MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD	
<small>The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.</small>	
	
Operator BENCH, JAMES	
Permit No 240032	
Date Issued 2/11/2014 Date Expires 2/11/2016	